



## Konformitätserklärung Declaration of conformity

**Hersteller / Manufacturer:** pfm medical mepro gmbh  
Am Söterberg 4, 66620 Nonnweiler, Germany

**SRN** DE-MF-000005190

Wir erklären in alleiniger Verantwortung, dass  
We declare under our sole responsibility that

**das Medizinprodukt / the medical device:** JetCan® Pro Safety Huber Needle

<b>REF:</b>	VJ01915S	VJ01919S	VJ01925S	VJ01938S	VJ02015S	VJ02019S
	VJ02025S	VJ02038S	VJ02215S	VJ02219S	VJ02225S	VJ02238S
	VJ61919S	VJ61925S	VJ61938S	VJ62019S	VJ62025S	VJ01919Y
	VJ01925Y	VJ01938Y	VJ02019Y	VJ02025Y	VJ02038Y	VJ02219Y
	VJ02225Y	VJ02238Y				

**Zweckbestimmung / Intended Use:** The JetCan® Pro Safety Huber Needle is indicated for use in the delivery of fluids and drugs, as well as blood sampling through surgically implanted vascular access ports.

**drugs, as well as blood sampling through surgically implanted vascular access ports.**

therapies, blood sampling and power injection of contrast media.

**Basic UDI-DI:** 4042301A0101ACFF

**Risikoklasse, Regel / Risk Class, Rule:** IIa, 7  
gemäß / according to MDR 2017/745, Annex VIII

die Anforderungen der Verordnung (EU) 2017/745 über Medizinprodukte erfüllt.  
meets the requirements of Regulation (EU) 2017/745 on medical devices.

**Konformitätsbewertungsverfahren gemäß:  
Conformity assessment procedure according to:** MDR (EU) 2017/745, Annex IX, Chapter II

**Angewandte gemeinsame Spezifikationen:  
Applied common specifications:** n/a

**Konformitätsbewertungsstelle / Notified Body:** DEKRA Certification GmbH  
Handwerkstr. 15, 70565 Stuttgart

**Kennnummer / Identification Number:** 0124

**Zertifikats-Nr. / Certificate No:** 51133-60-02-00

**Ort / Place:** Nonnweiler

**Datum / Date:** 15. Januar 2024

**Gültig bis / Expiry date:** 08.11.2025

**Unterschrift / Signature:**   
**Ausgestellt von / Issued by:** Christof Schmeer  
Vice President Quality & Regulatory Affairs  
pfm medical mepro gmbh